

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 07/06/2020

OPERATOR NAME: GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:24
INTERNAL STANDARD	VERIFIED	07:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:24
BLANK TEST	0.00	07:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:26
BLANK TEST	0.00	07:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28

