

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 07/08/2020

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:19
INTERNAL STANDARD	VERIFIED	08:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:22
BLANK TEST	0.00	08:23

RETURN FROM Calibration

