

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 07/12/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:46
INTERNAL STANDARD	VERIFIED	07:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:46
BLANK TEST	0.00	07:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:47
BLANK TEST	0.00	07:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:49
BLANK TEST	0.00	07:49

