

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 07/13/2020

OPERATOR NAME:
 GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:31
INTERNAL STANDARD	VERIFIED	07:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:31
BLANK TEST	0.00	07:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:32
BLANK TEST	0.00	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:33
BLANK TEST	0.00	07:34

