

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 07/22/2021

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:18
INTERNAL STANDARD	VERIFIED	10:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:20
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:22

