

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 07/27/2021

OPERATOR NAME: *JAA*
 OFC HINMAN
 Permit #: 39036
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:09
INTERNAL STANDARD	VERIFIED	11:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:09
BLANK TEST	0.00	11:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:10
BLANK TEST	0.00	11:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:11
BLANK TEST	0.00	11:12

JAA

