

SOLUTION CHANGE

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 08/03/2020

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:43
INTERNAL STANDARD	VERIFIED	15:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:46
BLANK TEST	0.00	15:47

