

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 08/03/2021

OPERATOR NAME:
 OFC HINMAN
 Permit #: 39036
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:47
INTERNAL STANDARD	VERIFIED	09:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:50
BLANK TEST	0.00	09:51

[Handwritten Signature] #0439

