

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 10/12/2021

OPERATOR NAME: GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:53
INTERNAL STANDARD	VERIFIED	07:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:54
BLANK TEST	0.00	07:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:55
BLANK TEST	0.00	07:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:56
BLANK TEST	0.00	07:57

