

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 10/13/2020

OPERATOR NAME: *[Signature]*
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:59
INTERNAL STANDARD	VERIFIED	08:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:01
BLANK TEST	0.00	09:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:02
BLANK TEST	0.00	09:03

