

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 10/18/2021

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30726  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:28
INTERNAL STANDARD	VERIFIED	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:29
BLANK TEST	0.00	07:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:31
BLANK TEST	0.00	07:32

