

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 11/08/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:00
INTERNAL STANDARD	VERIFIED	07:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04

