

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 11/09/2020

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:36
INTERNAL STANDARD	VERIFIED	08:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:38
BLANK TEST	0.00	08:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:39
BLANK TEST	0.00	08:40

