

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 11/29/2021

OPERATOR NAME: *GD*
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:52
INTERNAL STANDARD	VERIFIED	07:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:53
BLANK TEST	0.00	07:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:54
BLANK TEST	0.00	07:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:55
BLANK TEST	0.00	07:56

