

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 12/06/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:51
INTERNAL STANDARD	VERIFIED	07:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:52
BLANK TEST	0.00	07:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:53
BLANK TEST	0.00	07:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:54
BLANK TEST	0.00	07:55

