

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 12/07/2020

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:11
INTERNAL STANDARD	VERIFIED	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:11
BLANK TEST	0.00	10:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:12
BLANK TEST	0.00	10:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15

