

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 12/14/2020

OPERATOR NAME:
 GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:06
INTERNAL STANDARD	VERIFIED	12:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:07
BLANK TEST	0.00	12:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:08
BLANK TEST	0.00	12:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:09
BLANK TEST	0.00	12:10

