

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 12/28/2020

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:41
BLANK TEST	0.00	07:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:44
BLANK TEST	0.00	07:44

